



प.स. २०८१/८२

च.न. ८६५

प्रदेश सरकार
उद्योग, कृषि तथा मत्स्य विकास मन्त्रालय
पशुपन्छी तथा मत्स्य विकास निर्देशनालय

कोशी प्रदेश, विराटनगर, नेपाल

नेपाल संवत्: ११४५

डिप फ्रिज (Deep Freezer) खरिद गर्ने सम्बन्धी सूचना ।

प्रथम पटक प्रकाशित मिति: २०८२/०२/०८

यस निर्देशनालयको आ.व. ०८१/८२ को स्वीकृत कार्यक्रम अनुसार निम्न अनुसारको डिप फ्रिज (Deep Freezer) सार्वजनिक खरिद ऐन, २०६३ संशोधन समेतको दफा ८ को उपदफा (१) क (८) तथा सार्वजनिक खरिद नियमावली, २०६४ संशोधन समेत को नियम ३१ ख (१) बमोजिम त्यस्तो डिप फ्रिज (Deep Freezer) उत्पादक कम्पनी वा सो को आधिकारिक विक्रेताहरू बीच मात्र प्रतिस्पर्धा गराउने (क्याटलग सपिङ्ग) विधिबाट खरिद गर्नुपर्ने भएकोले सार्वजनिक खरिद नियमावली, २०६४ को नियम ३१ ख को उपनियम (२) अनुसार देहाय बमोजिमको डिप फ्रिज (Deep Freezer) विक्री गर्न चाहने उत्पादक वा आधिकारिक विक्रेताले आ-आफ्नो उत्पादनको आधिकारिक स्पेसिफिकेशन, गुणस्तर, मूल्य र सुविधा सहितको विवरण (क्याटलग वा ब्रोसर), डिप फ्रिज (Deep Freezer) उत्पादक कम्पनी वा सो को आधिकारिक विक्रेता प्रमाण पत्रको प्रमाणीत प्रतिलिपि, स्थायी लेखा (PAN) दर्ता प्रमाणपत्रको प्रमाणित प्रतिलिपि, आ.व. २०८०/८१ को कर चुक्ता प्रमाणपत्रको प्रमाणीत प्रतिलिपि, व्यवसाय संचालनको इजाजतपत्रको प्रमाणीत प्रतिलिपि संलग्न राखी यो सूचना प्रकाशन भएको मितिले १५ औं दिन, (मिति २०८२/०२/२२) दिनको १२.०० बजेसम्म यस निर्देशनालयमा सिलबन्दी प्रस्ताव दर्ता गर्नु हुन जानकारी गराइन्छ, दर्ता भएका प्रस्तावहरू सोही दिन दिनको २.०० बजे खोलिने हुँदा सम्बन्धित उत्पादक कम्पनी वा सो को आधिकारिक विक्रेताहरूको जानकारीका लागि यो सूचना प्रकाशित गरिएको छ । यस सम्बन्धि विस्तृत प्राविधिक स्पेसिफिकेशनको लागि यस निर्देशनालयमा सम्पर्क राख्नु हुन वा निर्देशनालयको वेबसाइट (www.dofdp.gov.np) बाट डाउनलोड गर्न सकिनेछ ।

डिप फ्रिज (Deep Freezer) को विवरण

Type of item	General Specification
Deep Freezer Refrigerator	Chest freezer hard top single or double door; Temperature range: +8 to -20 degree Celsius or below; Should be fit to store vaccines; Storage Capacity should not be less than 350 litres; Corrosion resistance; Chest type CFC with free insulation, Having at least 2 basket & door with lock and handle, CFC free Refrigerant; Compressor: ssealed heavy duty type that starts and operates on electricity supply from 165 to 255 volts, minimum 240 watt; Should be supplied with stabilizer with minimum 1 KVA (110 Volt to 250 Volt); Product offered should be designed to be stored and to operate normally under Nepal's Power Supply climate, temperature (0 °C to +45 °C), low and high humidity conditions; unit power supply: 220-240VAC, 50Hz fitted with appropriate plug & power cable minimum 3 metres long; Should comply with ISO certified product; Having comprehensive warranty for 2 years after acceptance.

अन्य विषयको हकमा स्वीकृत स्पेसिफिकेशन बमोजिम हुनेछ ।

पशुपन्छी तथा मत्स्य विकास निर्देशनालय,
कोशी प्रदेश, विराटनगर



2072/19/20
निदेशक

Technical Specification of Deep Freezer Refrigerator

S.N.	Particulars	DoLFD Biratnagar Requirements
1.	Description of Function	Chest freezer hard top single or double door.
2.	Temperature range:	+8 to -20 or below
3.	Operational Requirements	Should be fit to store vaccines.
4.	Storage Capacity	Not less than 350 litres
5.	Construction features:	<ul style="list-style-type: none">Corrosion resistance: Internal and external cabinet, lid and frame protected against corrosion. Bidder to specify the materials used for construction of internal & external cabinet.Chest type CFC – free insulationBasket -at least 2Door with lock and handle
6.	Refrigerant:	CFC free Refrigerant
7.	Compressor:	<ul style="list-style-type: none">Sealed heavy duty type.Starts and operates on electricity supply from 165 to 255 volts.Compressor should be minimum 240 watt.
8.	Stabilizer:	<ul style="list-style-type: none">Should be supplied with stabilizer.Minimum 1 KVA (110 Volt to 250 Volt)
9.	Operating Environment	<p>The product offered shall be designed to be stored and to operate normally under Nepal's Power Supply, Climate, Temperature (0 °C to +45 °C), low and high Humidity conditions.</p> <ul style="list-style-type: none">Unit Power supply: 220-240VAC, 50Hz fitted with appropriate plug. The power cable must be minimum 3 metres long.
10.	Standards and Safety Requirements	<ul style="list-style-type: none">Shall comply with ISO certified product.
11.	Warranty	<ul style="list-style-type: none">Comprehensive warranty for 2 years after acceptance.
12.	Maintenance Service During Warranty Period	<ul style="list-style-type: none">During the warranty period supplier must ensure corrective/breakdown maintenance whenever required.
13.	Installation and Commissioning	<ul style="list-style-type: none">The bidder must arrange for the equipment to be installed by certified or qualified personnel and should train the DoLFD staffs about operation of the Freeze.
14.	Documentation	<ul style="list-style-type: none">User (Operating) manual in English.
15.	Delivery and testing	<ul style="list-style-type: none">The bidder must arrange the delivery of goods to Directorate of Livestock & Fisheries Development (DoLFD) office and should assure the complete testing of the goods in presence of a mechanical engineer.

2072/19/20
निदेशक