



प.स. २०८१/८२
च.न.

प्रदेश सरकार
उद्योग, कृषि तथा सहकारी मन्त्रालय

पशुपन्छी तथा मत्स्य विकास निर्देशनालय

कोशी प्रदेश, विराटनगर, नेपाल



नेपाल संवत्: १९४५

डिप फ्रिज (Deep Freezer) खरिद गर्ने सम्बन्धी सूचना ।

दोस्रो पटक प्रकाशित मिति: २०८२/०२/२७

यस निर्देशनालयको आ.व. ०८१/८२ को स्वीकृत कार्यक्रम अनुसार निम्न अनुसारको २५ (पच्चीस) थान डिप फ्रिज (Deep Freezer) सार्वजनिक खरिद ऐन, २०६३ संशोधन समेतको दफा ८ को उपदफा (१) क (८) तथा सार्वजनिक खरिद नियमावली, २०६४ संशोधन समेत को नियम ३१ ख (१) बमोजिम त्यस्तो डिप फ्रिज (Deep Freezer) उत्पादक कम्पनी वा सो को आधिकारिक विक्रेताहरू बीच मात्र प्रतिस्पर्धा गराउने (क्याटलग सपिङ्ग) विधिबाट खरिद गर्नुपर्ने भएकोले सार्वजनिक खरिद नियमावली, २०६४ को नियम ३१ ख को उपनियम (२) अनुसार देहाय बमोजिमको डिप फ्रिज (Deep Freezer) विक्री गर्न चाहने उत्पादक वा आधिकारिक विक्रेताले तपसिलका कागजातहरू संलग्न राखि यो सूचना प्रकाशन भएको मितिले ७ औं दिन, दिनको १२.०० बजेसम्म यस निर्देशनालयमा सिलबन्दी प्रस्ताव दर्ता गर्नु हुन जानकारी गराइन्छ, दर्ता भएका प्रस्तावहरू सोही दिन दिनको २.०० बजे खोलिने हुँदा सम्बन्धित उत्पादक कम्पनी वा सो को आधिकारिक विक्रेताहरूको जानकारीका लागि यो सूचना प्रकाशित गरिएको छ । यस सम्बन्धित विस्तृत प्राविधिक स्पेसिफिकेशनको लागि यस निर्देशनालयमा सम्पर्क राख्नु हुन वा निर्देशनालयको वेबसाइट (www.doldf.pl.gov.np) बाट डाउनलोड गर्न सकिनेछ ।

तपसिल:

- (१) आ-आफ्नो उत्पादनको आधिकारिक स्पेसिफिकेशन, गुणस्तर, मूल्य (ने.रु. भ्याट सहित) र सुविधा सहितको विवरण (प्रस्ताव पत्रका साथै क्याटलग वा ब्रोसर, वेबसाइटमा खुल्लुपर्ने; स्पेसिफिकेशन अनुसारको गुणस्तर प्रमाणित हुने कागजात संलग्न हुनुपर्ने) ।
- (२) डिप फ्रिज (Deep Freezer) उत्पादक कम्पनी वा सो को आधिकारिक विक्रेता प्रमाण पत्रको प्रमाणीत प्रतिलिपि ।
- (३) व्यवसाय संचालनको इजाजतपत्रको प्रमाणीत प्रतिलिपि ।
- (४) स्थायी लेखा (PAN) दर्ता प्रमाणपत्रको प्रमाणीत प्रतिलिपि ।
- (५) आ.व. २०८०/८१ को कर चुक्ता प्रमाणपत्रको प्रमाणीत प्रतिलिपि ।

डिप फ्रिज (Deep Freezer) को विवरण

Type of item	General Specification
Deep Freezer Refrigerator (No. of items/ Quantity:- 25)	Chest freezer hard top single or double door; Temperature range: +8 to -20 or below; Should be fit to store vaccines; Storage Capacity should not be less than 350 litres; Corrosion resistance; Chest type CFC – free insulation, Having at least 2 basket, door with lock and handle, CFC free Refrigerant; Compressor: sealed heavy duty type that starts and operates on electricity supply from 165 to 255 volts, minimum 240 watt; Should be supplied with stabilizer with minimum 1 KVA (110 Volt to 250 Volt); Product offered should be designed to be stored and to operate normally under Nepal's Power Supply climate, temperature (0 °C to +45 °C), low and high humidity conditions; unit power supply: 220-240VAC, 50Hz fitted with appropriate plug & power cable minimum 3 metres long; Should comply with ISO certified product; Having comprehensive warranty for 2 years after acceptance.

अन्य विषयको हकमा स्वीकृत स्पेसिफिकेशन बमोजिम हुनेछ ।

पशुपन्छी तथा मत्स्य विकास निर्देशनालय,
कोशी प्रदेश, विराटनगर

(दोस्रो पटक प्रकाशन मिति : २०८२।०२।२७ गते)

यस नगरपालिकाको वार्षिक स्वीकृत कार्यक्रम अन्तर्गत E-Bidding मार्फत तपशिल बमोजिमको सार्वजनिक खरिद कार्य गर्नुपर्ने भएकोले तपसिल कार्य गर्न इच्छुक र इजाजत प्राप्त फर्म/कम्पनी/संस्थाले रितपूर्वकको बोलपत्र पेश गर्नुहुन यो सूचना प्रकाशित गरिएको छ ।

तपशिल:

ठेक्का नं.	विवरण	बोलपत्र फारम दस्तुर	जमानत रकम	खाता विवरण
TBM/G/NCB-34/081/82	Supply, Delivery, Installation and Testing of Organic Fertilizer Production Machine	३,०००/-	६०,०००/-	१) आन्तरिक राजस्व खाता: १६१२०५००२८२९६००२ २) धरोटी खाता: १६१२०५०५२३८८८०१३ बैंक: NIC Asia Bank, Bhurigaun Branch

यो सूचनाको विस्तृत जानकारी का लागि PPMO को Website: www.bolpatra.gov.np/egg बाट Download गरि प्राप्त गर्न सकिने छ वा यस नगरपालिकाको कार्यालयमा कार्यालय समयभित्र सम्पर्क गर्न सकिने छ ।



प्रदेश सरकार
उद्योग, कृषि तथा सहकारी मन्त्रालय
पशुपन्थी तथा मत्स्य विकास निर्देशनालय
कोशी प्रदेश, विराटनगर, नेपाल

डिप फ्रिज (Deep Freezer) खरिद गर्ने सम्बन्धी सूचना

दोस्रो पटक प्रकाशित मिति: २०८२/०२/२७

यस निर्देशनालयको आ.ब. ०८१/८२ को स्वीकृत कार्यक्रम अनुसार निम्न अनुसारको २५ (पच्चीस) थान डिप फ्रिज (Deep Freezer) सार्वजनिक खरिद ऐन, २०६३ संशोधन समेतको दफा ८ को उपदफा (१) क (८) तथा सार्वजनिक खरिद नियमावली, २०६४ संशोधन समेत को नियम ३१ ख (१) बमोजिम त्यस्तो डिप फ्रिज (Deep Freezer) उत्पादक कम्पनी वा सो को आधिकारिक विक्रेताहरु बीच मात्र प्रतिस्पर्धा गराउने (क्वाटलग सपिङ) विधिबाट खरिद गर्नुपर्ने भएकोले सार्वजनिक खरिद नियमावली, २०६४ को नियम ३१ ख को उपनियम (२) अनुसार देहाय बमोजिमको डिप फ्रिज (Deep Freezer) विक्री गर्न चाहने उत्पादक वा आधिकारिक विक्रेताले तपसिलका कागजातहरु संलग्न राखि यो सूचना प्रकाशन भएको मितिले ७ औं दिन, दिनको १२.०० बजेसम्म यस निर्देशनालयमा सिलबन्दी प्रस्ताव दर्ता गर्नु हुन जानकारी गराइन्छ, दर्ता भएका प्रस्तावहरु सोही दिन दिनको २.०० बजे खोलीने हुँदा सम्बन्धित उत्पादक कम्पनी वा सो को आधिकारिक विक्रेताहरुको जानकारीका लागि यो सूचना प्रकाशित गरिएको छ । यस सम्बन्धि विस्तृत प्राविधिक स्पेसिफिकेशनको लागि यस निर्देशनालयमा सम्पर्क राख्नु हुन वा निर्देशनालयको वेबसाइट (www.dofid.pl.gov.np) बाट डाउनलोड गर्न सकिनेछ ।

तपसिल:

- (१) आ-आफ्नो उत्पादनको आधिकारिक स्पेसिफिकेशन, गुणस्तर, मूल्य (ने.रु. भ्याट सहित) र सुविधा सहितको विवरण (प्रस्ताव पत्रका साथै क्वाटलग वा क्रोसर, वेबसाइटमा खुल्नुपर्ने, स्पेसिफिकेशन अनुसारको गुणस्तर प्रमाणित हुने कागजात संलग्न हुनुपर्ने) ।
- (२) डिप फ्रिज (Deep Freezer) उत्पादक कम्पनी वा सो को आधिकारिक विक्रेता प्रमाण पत्रको प्रमाणीत प्रतिलिपि ।
- (३) व्यवसाय संचालनको इजाजतपत्रको प्रमाणीत प्रतिलिपि ।
- (४) स्थायी लेखा (PAN) दर्ता प्रमाणपत्रको प्रमाणीत प्रतिलिपि ।
- (५) आ.ब. २०८०/८१ को कर चुक्ता प्रमाणपत्रको प्रमाणीत प्रतिलिपि ।

डिप फ्रिज (Deep Freezer) को विवरण

Type of item	General Specification
Deep Freezer Refrigerator (No. of items/Quantity:- 25)	Chest freezer hard top single or double door; Temperature range: +8 to -20 or below; Should be fit to store vaccines ; Storage Capacity should not be less than 350 litres; Corrosion resistance; Chest type CFC – free insulation, Having at least 2 basket, door with lock and handle, CFC free Refrigerant; Compressor: sealed heavy duty type that starts and operates on electricity supply from 165 to 255 volts, minimum 240 watt; Should be supplied with stabilizer with minimum 1 KVA (110 Volt to 250 Volt); Product offered should be designed to be stored and to operate normally under Nepal's Power Supply climate, temperature(0 °C to +45 °C), low and high humidity conditions; unit power supply: 220/240VAC, 50Hz fitted with appropriate plug & power cable minimum 3 metres long; Should comply with ISO certified product; Having comprehensive warranty for 2 years after acceptance.

अन्य विषयको हकमा स्वीकृत स्पेसिफिकेशन बमोजिम हुनेछ ।

पशुपन्थी तथा मत्स्य विकास निर्देशनालय,
कोशी प्रदेश, विराटनगर

डिभिजन वन कार्यालय भक्तपुरको पञ्चवर्षिय/रणनीतिक योजना (आ.व. २०८२/०८३-२०८६/८७) कार्यान्वयनको वातावरणीय अध्ययन प्रतिवेदन तयारी सम्बन्धी सार्वजनिक सुनुवाई कार्यक्रम र राय सुभाषको लागि सूचना



2072/19/20
निदेशक

Technical Specification of Deep Freezer Refrigerator

S.N.	Particulars	DoLFD Biratnagar Requirements
1.	Description of Function	Chest freezer hard top single or double door.
2.	Temperature range:	+8 to -20 or below
3.	Operational Requirements	Should be fit to store vaccines.
4.	Storage Capacity	Not less than 350 litres
5.	Construction features:	<ul style="list-style-type: none">▪ Corrosion resistance: Internal and external cabinet, lid and frame protected against corrosion. Bidder to specify the materials used for construction of internal & external cabinet.▪ Chest type CFC – free insulation▪ Basket -at least 2▪ Door with lock and handle
6.	Refrigerant:	CFC free Refrigerant
7.	Compressor:	<ul style="list-style-type: none">▪ Sealed heavy duty type.▪ Starts and operates on electricity supply from 165 to 255 volts.▪ Compressor should be minimum 240 watt.
8.	Stabilizer:	<ul style="list-style-type: none">▪ Should be supplied with stabilizer.▪ Minimum 1 KVA (110 Volt to 250 Volt)
9.	Operating Environment	<p>The product offered shall be designed to be stored and to operate normally under Nepal's Power Supply, Climate, Temperature (0 °C to +45 °C), low and high Humidity conditions.</p> <ul style="list-style-type: none">▪ Unit Power supply: 220-240VAC, 50Hz fitted with appropriate plug. The power cable must be minimum 3 metres long.
10.	Standards and Safety Requirements	<ul style="list-style-type: none">▪ Shall comply with ISO certified product.
11.	Warranty	<ul style="list-style-type: none">▪ Comprehensive warranty for 2 years after acceptance.
12.	Maintenance Service During Warranty Period	<ul style="list-style-type: none">▪ During the warranty period supplier must ensure corrective/breakdown maintenance whenever required.
13.	Installation and Commissioning	<ul style="list-style-type: none">▪ The bidder must arrange for the equipment to be installed by certified or qualified personnel and should train the DoLFD staffs about operation of the Freeze.
14.	Documentation	<ul style="list-style-type: none">▪ User (Operating) manual in English.
15.	Delivery and testing	<ul style="list-style-type: none">▪ The bidder must arrange the delivery of goods to Directorate of Livestock & Fisheries Development (DoLFD) office and should assure the complete testing of the goods in presence of a mechanical engineer.

2072/19/20
निदेशक

(Format to be filled by bidder submitting technical proposal)

Bidder's Offer on supply of Deep Freezer Refrigerator

S.N.	Particulars	DoLFD Biratnagar Requirements	Bidder's Offer (Mention Comply/Not comply and details of the deep freezer to be supplied for each particulars that should be verified by supporting documents)	Remarks
1.	Description of Function	Chest freezer hard top single or double door.		
2.	Temperature range:	+8 to -20 or below		
3.	Operational Requirements	Should be fit to store vaccines.		
4.	Storage Capacity	Not less than 350 litres		
5.	Construction features:	<ul style="list-style-type: none"> Corrosion resistance: Internal and external cabinet, lid and frame protected against corrosion. Bidder to specify the materials used for construction of internal & external cabinet. Chest type CFC – free insulation Basket -at least 2 Door with lock and handle 		
6.	Refrigerant:	CFC free Refrigerant		
7.	Compressor:	<ul style="list-style-type: none"> Sealed heavy duty type. Starts and operates on electricity supply from 165 to 255 volts. Compressor should be minimum 240 watt. 		
8.	Stabilizer:	<ul style="list-style-type: none"> Should be supplied with stabilizer. Minimum 1 KVA (110 Volt to 250 Volt) 		
9.	Operating Environment	The product offered shall be designed to be stored and to operate normally under Nepal's Power Supply, Climate, Temperature (0 °C to +45 °C), low and high Humidity conditions.		
		<ul style="list-style-type: none"> Unit Power supply: 220-240VAC, 50Hz fitted with appropriate plug. The power cable must be minimum 3 metres long. 		
10.	Standards and Safety Requirements	<ul style="list-style-type: none"> Shall comply with ISO certified product. 		

11.	Warranty	<ul style="list-style-type: none"> Comprehensive warranty for 2 years after acceptance. 		
12.	Maintenance Service During Warranty Period	<ul style="list-style-type: none"> During the warranty period supplier must ensure corrective/breakdown maintenance whenever required. 		
13.	Installation and Commissioning	<ul style="list-style-type: none"> The bidder must arrange for the equipment to be installed by certified or qualified personnel and should train the DoLFD staffs about operation of the Freeze. 		
14.	Documentation	<ul style="list-style-type: none"> User (Operating) manual in English. 		
15.	Delivery and testing	<ul style="list-style-type: none"> The bidder must arrange the delivery of goods to Directorate of Livestock & Fisheries Development (DoLFD) office and should assure the complete testing of the goods in presence of a mechanical engineer. 		